

Calibration Certificate

Issued By: Castle Group Ltd

Date Of Issue : 14/06/17

Certificate No : 000000/00000

Page 1 of 3



All instruments are tested to check compliance with particular specifications. These specifications may be appropriate British Standards, or if the instrument was not originally designed to meet any British Standard, or when the instrument was originally manufactured a relevant British Standard did not exist, the instrument will be tested to the manufacturer's original specification.

Absolute calibration of vibration calibrators and vibration meters is checked at one or more standard frequencies against an independent sound source with calibration directly traceable to the Physikalisch-Technische Bundesanstalt (PTB) in Germany. The PTB reference applicable for the calibration of the test equipment is shown below.

The performance of the instrument was determined by comparison with the manufacturers' specification as found in the instrument handbook or other technical publication. Any significant uncertainty of the measuring system will also be included.

The instrument was allowed to stabilise for a period of 30 minutes prior to measurements made.

The ambient temperature and relative humidity throughout calibration were 22 ± 2 °C and 50% RH respectively.

Instruments used to carry out this calibration are as follows: -

Vibration Calibrator: SV111 Serial Number 25039

Vibration Exciter: GA606 Serial Number 4357

Applicable Reference: PTB-17012-2017, reference no: 1.71-1508/0317.

Subject of Calibration: GA2005H

Instrument: HARM (Hand Arm Risk Measurement) Vibration Meter

Serial No: 000000

Accelerometer Type: KD1006

X Axis:

Sensitivity (mV/g) Y Axis:

Z Axis:

Accelerometer Serial No:

X Axis:

Calibration Coefficient Y Axis:

Z Axis:

Basis Of Test: Compliance to Manufacturer's Original Specification

Calibrated By:

A N Engineer

(Approved Signatory)

Date of Calibration: 14 Jun 2017

Completed Status: Pass

Checked By:

A N O Engineer

(Approved Signatory)

Recalibration Due: 01 Jul 2018

Client: End User Company Name

Address: Place

Street

Location

Town

County

Postcode

Client Reference: If Applicable

Castle Group Ltd

Salter Road, Scarborough Business Park, Scarborough, North Yorkshire YO11 3UZ United Kingdom

t: +44 (0)1723 584250 f: +44 (0)1723 583728 e: sales@castlegroup.co.uk

www.castlegroup.co.uk

CALIBRATION RESULTS

Page 2 of 3 Pages

Date: 14/06/17 Certificate No: 000000/00000 Calibrated By: A N Engineer



INDICATED RESPONSE (Reference Range) Relative to 10.02ms² at 159.20Hz

HARM (TRANSDUCER = 10mV)

| PARAMETER | ACTUAL VALUE | | | PREFERRED VALUE | TOLERANCE |
|------------------------------|--------------|---|---|-----------------|--------------------|
| | X | Y | Z | | |
| METRIC (ms ²) | | | | 1.009 | 0.9525 to 1.0687 |
| dB (Ref 1e-6) | | | | 120.077 | 119.577 to 120.577 |
| IMPERIAL (fts ²) | | | | 3.310 | 3.125 to 3.506 |
| g | | | | 0.1029 | 0.0971 to 0.1090 |

RANGE TEST

HARM Weighted Response - Relative to 10.0ms² at 160Hz

LINEARITY RESPONSE The method employed was a direct electrical signal injection on the reference range.

| TARGET LEVEL (ms ²) | READING | | | TOLERANCE (ms ²) |
|---------------------------------|---------|---|---|------------------------------|
| | X | Y | Z | |
| 1.00 | | | | 0.95 - 1.05 |
| 10.00 | | | | 9.50 - 10.50 |
| 100.00 | | | | 95.00 - 105.00 |

LINEARITY RESPONSE (Other Measurement Ranges) The method employed was a direct electrical signal injection.

| RANGE | TARGET LEVEL (ms ²) | READING (X ONLY) | TOLERANCE (ms ²) |
|---------------|---------------------------------|------------------|------------------------------|
| 0.02 - 100.0 | 0.10 | | 0.095 - 0.105 |
| 0.02 - 100.0 | 1.00 | | 0.95 - 1.05 |
| 0.02 - 100.0 | 10.00 | | 9.50 - 10.50 |
| 0.1 - 500.0 | 0.10 | | 0.095 - 0.105 |
| 0.1 - 500.0 | 1.00 | | 0.95 - 1.05 |
| 0.1 - 500.0 | 10.00 | | 9.50 - 10.50 |
| 1.00 - 5000.0 | 1.00 | | 0.95 - 1.05 |
| 1.00 - 5000.0 | 10.00 | | 9.50 - 10.50 |
| 1.00 - 5000.0 | 100.00 | | 95.00 - 105.00 |

Castle Group Ltd

Salter Road, Scarborough Business Park, Scarborough, North Yorkshire YO11 3UZ United Kingdom
t: +44 (0)1723 584250 f: +44 (0)1723 583728 e: sales@castlegroup.co.uk

www.castlegroup.co.uk

CALIBRATION RESULTS

Page 3 of 3 Pages

Date: 14/06/17 Certificate No: 000000/00000 Calibrated By: A N Engineer



FILTER TESTS

HARM (Hand Arm Risk Measurement)

Wh FILTER TEST. The method employed was a direct electrical signal injection on the reference range.

| FREQUENCY (Hz) | TARGET LEVEL (ms ⁻²) | READING | | | TOLERANCE (ms ⁻²) |
|-------------------|-------------------------------------|---------|---|---|----------------------------------|
| | | X | Y | Z | |
| 2.0 | 9.92 | | | | 0.00 - 12.06 |
| 5.0 | 54.14 | | | | 44.51 - 65.84 |
| 16.0 | 89.92 | | | | 82.04 - 96.38 |
| 80.0 | 20.09 | | | | 18.54 - 21.78 |
| 160.0 | 10.00 | | | | 9.95 - 10.05 |
| 800.0 | 1.85 | | | | 1.71 - 2.00 |
| 1,600.0 | 0.53 | | | | 0.44 - 0.65 |

Uncertainties of the measurement

The uncertainties in the table of results correspond to an estimated confidence probability of not less than 95%.

* *Notes*

Castle Group Ltd

Salter Road, Scarborough Business Park, Scarborough, North Yorkshire YO11 3UZ United Kingdom
t: +44 (0)1723 584250 f: +44 (0)1723 583728 e: sales@castlegroup.co.uk

www.castlegroup.co.uk